



EU DECLARATION OF CONFORMITY

issued under the sole responsibility of the manufacturer

According to Regulation (EU) 2017/745 of the European Parliament and of the Council on medical devices,

Manufacturer: Kettex Development sro
Registered office: Na pěšině 465/6, Prague 8, 184 00, Czech Republic
ID: 3267016
Actor ID/Germany: CZ-MF-000004085

Product identification data

Name: FULL HD ENDOSCOPY CAM KX-02_Kymo
Base UDI-DI: 8594200760CAMKX-KymoGR

The manufacturer **declares** under its sole responsibility that the characteristics of the above-mentioned medical device meet all the requirements in accordance with Regulation (EU) 2017/745 of the European Parliament and of the Council, and that this medical device is safe, effective and suitable for the provision of healthcare for its intended purpose. The manufacturer further declares that it has taken measures to ensure the conformity of the medical device placed on the market with the essential requirements and the manufacturer's technical documentation.

Intended use:

The intended purpose of the Full HD Endoscopy Cam KX-02_Kymo video endoscopic medical device is digitization the image visible in the eyepiece of the endoscope and its subsequent display on the PC. The medical device allows image manipulation to improve contrast. Medical device is intended for outpatient care.

Risk class: I, non-sterile, non-measuring

Standards used : ČSN EN ISO 20417:2021, ČSN EN ISO 14971:2020/A11:2022, ČSN EN ISO 15223-1:2022, ČSN EN ISO 10993-1:2021, ČSN EN 60601-1:2007 +A11:2012 +A1:2014 + A12:2015 + A2:2022 + A13:2025 , ČSN EN 60601-1-2:2016/A1:2021, ČSN EN 60601-2-18 ed. 2:2016, ČSN EN 62366-1:2019/A1:2021, ČSN EN 60601-1-6 ed. 3:2010 + A1:2015/A2:2021

In Prague on January 26, 2026

RNDr. Filip Krolupper, PhD
Managing Director of Kettex Development sro